




Ref. Certif. No.

**DK-151580-UL**

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**

**CB TEST CERTIFICATE**

Product	AC/DC Power Module
Name and address of the applicant	MINMAX Technology Co. Ltd. 18 Sin-Sin Road, An-Ping Industrial District, Tainan City, 702 Taiwan
Name and address of the manufacturer	MINMAX Technology Co. Ltd. 18 Sin-Sin Road, An-Ping Industrial District, Tainan City, 702 Taiwan
Name and address of the factory	MINMAX Technology Co. Ltd. 18 Sin-Sin Road, An-Ping Industrial District, Tainan City, 702 Taiwan
Note: When more than one factory, please report on page 2	<input type="checkbox"/> Additional Information on page 2
Ratings and principal characteristics	Input: 100-240Vac, 60/50 Hz, 0.6A-0.3A <input checked="" type="checkbox"/> Additional Information on page 2
Trademark / Brand (if any)	
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	AJM-24SaX, AJM-24DbX <input checked="" type="checkbox"/> Additional Information on page 2
Additional information (if necessary may also be reported on page 2)	National Differences: CA, JP, US <input type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020
As shown in the Test Report Ref. No. which forms part of this Certificate	2309053-CB issued on 2024-03-25

This CB Test Certificate is issued by the National Certification Body



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2024-03-26

Signature:  
Thomas Wilson



Ref. Certif. No.

**DK-151580-UL**

**Additional Model Detail(s):**

AJM-24SaX, AJM-24DbX, (where a can be 05, 09, 12, 15, 24. b can be 12, 15. X can be C or blank)

**Additional Ratings:**

Output:

Models Output

AJM-24S05X 5.0 V dc, 3000 mA

AJM-24S09X 9.0 V dc, 2666 mA

AJM-24S12X 12 V dc, 2000 mA

AJM-24S15X 15 V dc, 1600 mA

AJM-24S24X 24 V dc, 1000 mA

AJM-24D12X +/-12 V dc, +/- 1000 mA

AJM-24D15X +/-15 V dc, +/- 800 mA

**Additional information (if necessary)**



- UL Solutions (US), 333 Pflugsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2024-03-26

Signature:

Thomas Wilson