



Ref. Certif. No.

DK-151792-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	AC/DC Power Module
Name and address of the applicant	MINMAX Technology Co. Ltd. 18 Sin-Sin Road, An-Ping Industrial District, Tainan City, 702 Taiwan.
Name and address of the manufacturer	MINMAX Technology Co. Ltd. 18 Sin-Sin Road, An-Ping Industrial District, Tainan City, 702 Taiwan.
Name and address of the factory	MINMAX Technology Co. Ltd. 18 Sin-Sin Road, An-Ping Industrial District, Tainan City, 702 Taiwan
Note: When more than one factory, please report on page 2	<input type="checkbox"/> Additional Information on page 2
Ratings and principal characteristics	Input: 100-240Vac, 60/50 Hz, 1.1-0.6 A <input checked="" type="checkbox"/> Additional Information on page 2
Trademark / Brand (if any)	
Customer's Testing Facility (CTF) Stage used	
Model / Type Ref.	APM-40SaX, APM-40DbX (where a can be 05, 12, 15, 24; b can be 12, 15; X can be C or blank)
Additional information (if necessary may also be reported on page 2)	<b>Additionally evaluated to:</b> EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021 National Differences: CA, JP, US <input type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020
As shown in the Test Report Ref. No. which forms part of this Certificate	2309054-CB issued on 2024-03-27

This CB Test Certificate is issued by the National Certification Body



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2024-04-01

Signature: Thomas Wilson



Ref. Certif. No.

**DK-151792-UL**

**Additional Ratings:**

Output:

Models	Output
APM-40S05X	5.0 V dc, 8000 mA
APM-40S12X	12 V dc, 3330 mA
APM-40S15X	15 V dc, 2660 mA
APM-40S24X	24 V dc, 1660 mA
APM-40D12X	+/-12 V dc, +/- 1660 mA
APM-40D15X	+/-15 V dc, +/- 1330 mA

**Additional information (if necessary)**



- UL Solutions (US), 333 Pflingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2024-04-01

Signature:

Thomas Wilson